

Tribo[™] Chloramphenicol Test Strip (Catalog No. TBS11121)

General Description

Chloramphenicol is a widely used broad-spectrum antibiotic effective against both Gram-positive and Gram-negative bacteria, and functions by inhibiting bacterial protein synthesis. It has been used to s could result from its usage. Its major toxic side effect is bone marrow suppression, lowering counts of white blood cells and platelets and sometime leading to fatal aplastic anemia. In addition, residual chloramphenicol can lead to antibiotic-resistant bacterial population. As a result, usage of chloramphenicol has been forbidden in many countries, and no detectable amount is allowed in any food of animal origin. Still, because of its low cost and high availability, chloramphenicol is still used illegally in many parts of the world.

The TriboTM chloramphenicol test strip provides a rapid and convenient test for chloramphenicol with a colloidal gold immunochromatographic design. It detects residual chloramphenicol in animal tissue such as meat, fish, or shrimp, providing a **fast** (results shown in 10 minutes), **simple**, **sensitive** and **reliable** detection approach for the presence of chloramphenicol. The lower detection limit (LDL) of chloramphenicol in the sample is 0.1 ng/ml (0.1 ppb).

Intended Use

The Tribo[™] chloramphenicol test strip is a lateral flow strip test for the detection of residual chloramphenicol in animal tissue (meat, fish, or shrimp). Assay sensitivity: 0.1 ng/ml (0.1 ppb).

Safety Instructions

To receive complete safety information on this product, contact Tribioscience, Inc. and request Material Safety Data Sheets.

Assay Principle

TriboTM chloramphenicol test strip is based on the principle of colloidal gold immunochromatography. An anti-chloramphenicol antibody is conjugated to colloidal gold and placed on conjugate pad. Colloidal gold provides red color to visualize antibody-antigen binding. Chloramphenicol antigen is immobilized on nitrocellulose membrane. After test sample is loaded onto sample pad, it mixes with gold-antibody conjugate and migrates along the membrane. If the sample contains no chloramphenicol or amount of chloramphenicol lower than detection limit, antibody conjugated to colloidal gold will bind the antigen immobilized on membrane, leading to clear red color presented on membrane detection lime where the test antigen is immobilized. If chloramphenicol concentration is higher than detection limit in the test sample, it will bind gold-antibody conjugate and prevent its binding onto the antigen line on membrane. As a result, no color will be visible on detection line on



membrane.

Reagents and materials in each pack

- a) 1 chloramphenicol test strip.
- b) 1 dropper tube (disposable).
- c) 1 pack of desiccant.
- d) 1 bottle of sample buffer.

Sample Collection and Test Procedure

- Equilibrate test strip to room temperature (20-25°C)
- ★ Equilibrate test sample to room temperature (20-25°C)
- a) Cut 1 g of tissue sample into tiny pieces and put into a centrifuge tube.
- b) Add 1 ml sample buffer, tighten the lid, and heat 5-10 min at 90°C (heating with water bath is recommended) before taking the tube out.
- c) Centrifuge at 4,000 rpm for 10 min.
- d) Cool down to room temperature, and add 2 drops of supernatant into the sample loading well of test strip and start timing. Observe results in 5 min.


Result Interpretation

Test result is interpreted by observing test line and control line shown in result window. Negative

(-): both test and control lines are present, and test line is darker than control line, or with darkness equal to control line. This demonstrates that chloramphenicol concentration in sample is lower than 0.1 ng/ml;

Positive (+): control line is present, and test line is either absent or lighter than control line. This result indicates chloramphenicol centration of higher than 0.1 ng/ml in sample;

Invalid test: no control line is present. Please repeat the test using a new test strip following instructions on this user guide.



Performance Data

- 1. Test strip is for one-time use only. Please use the test strip on the same day the package is opened.
- 2. Do not use tap water, purified water, or distilled water as negative control.



- 3. Test should be performed at room temperature, and test strip and sample need to be equilibrated to room temperature before the test.
- 4. If no liquid movement is observed in the test window 30 seconds after test solution (supernatant) is added to sample loading well, add one more drop of test solution.
- 5. It is recommended that positive samples identified by test strip be re-tested with more accurate approaches (such as HPLC or GC-MS).

Storage and Expiration Date

Storage: Room Temperature.

Expiration Date: 12 months after manufacturing date.

Technical Assistance

For ordering or technical assistance regarding this kit, or for additional information about TribioScience products, please email info@tribioscience.com or call (650) 917-9269.

General Limited Warranty

TribioScience, Inc. warrants its manufactured products against defects in materials and workmanship when used in accordance with the applicable instructions for a period not to extend beyond a product's printed expiration date. TribioScience makes no other warranty, expressed or implied. There is no warranty of merchantability or fitness for a particular purpose. The warranty provided herein and the data, specifications and descriptions of TribioScience products appearing in published catalogues and product literature may not be altered except by express written agreement signed by an officer of TribioScience. Representations, oral or written, which are inconsistent with this warranty or such publications are not authorized and, if given, should not be relied upon.

In the event of a breach of the foregoing warranty, TribioScience Inc.'s sole obligation shall be to repair or replace, at its option, any product or part thereof that proves defective in materials or workmanship within the warranty period, provided the customer notifies TribioScience promptly of any such defect. The exclusive remedy provided herein shall not be deemed to have failed of its essential purpose so long as TribioScience is willing and able to repair or replace any nonconforming TribioScience product or part. TribioScience shall not be liable for consequential, incidental, special or any other indirect damages resulting from economic loss or property damage sustained by a customer from the use of its products. However, in some states the purchaser may have rights under state law in addition to those provided by this warranty.

For research use only.