

For the quantitative determination concentrations of mouse KC in cell culture supernatants, serum and plasma.

INTRODUCTION


Keratinocyte chemoattractant (KC), also known as CXCL1, SCYB1, or N51 is a member of the CXC or alpha chemokine family. KC is similar in structure and activity to IL-8 and MIP 2. It is produced by melanocytes, epidermal keratinocytes, monocytes, macrophages, mammary epithelial cells, endothelial cells, neutrophils, fibroblasts and hepatocytes. Expression level of this chemokine is normally very low in most tissues however, it can be significantly elevated in many disease states including rheumatoid arthritis, cancer, and experimental autoimmune encephalopathy, post surgery and other inflammatory responses.

Tribioscience’s Fast Mouse KC ELISA is designed to quantitatively detect mouse KC levels in different tissues including skin, muscle, neural, serum, and other biological samples. The main feature is that **the kit uses our novel proprietary approaches to combine samples and detections into a one-step instead of the complicated traditional methods. It makes the assay simple, easy, accurate, and fast. The measurement can be finished in 2 hours, with no need for 4-5 hours (Fig. 1). The detection range is from 15 to 1000 pg/mL.** The levels of mouse KC samples are parallel to the standard curves obtained using the kit standards linearly. These results indicate that this kit can be used to determine relative mass values for natural mouse KC protein.

PRINCIPLE OF THE ASSAY

This assay employs our novel proprietary sandwich enzyme immunoassay techniques (see Fig. 1). A monoclonal antibody specific for mouse KC is pre-coated onto a microplate. Standards or samples and a biotin-conjugated detection antibody are pipetted into the wells and concurrently incubated to form a sandwich complex in one step. Simply aspirate each well without washing and directly add Streptavidin-HRP into the complex. Following a wash, an **ultra-sensitive TMB substrate solution** is added to the wells for color development. The color intensity is proportional to the amount of KC bound in the initial step. The intensity of the color is measured by plate reading at 450 nm.

Fig. 1



KIT CONTENT AND STORAGE CONDITIONS

PART	PART#	DESCRIPTION	STORAGE OF OPENED/ RECONSTITUTED
Mouse KC Microplate	TBS3060A	96 well polystyrene microplate (12 strips of 8 wells) coated with a polyclonal antibody specific for mouse KC.	Return unused wells to the foil pouch. Reseal along entire edge of the zip-seal. May be stored for up to 1 month at 2-8 °C.
Mouse KC Standard	TBS3060B	30 µL of Recombinant mouse KC protein (50 ng/mL).	Aliquot and store at -20 °C for up to 1 month in a manual defrost freezer. Avoid repeated freeze-thaw cycles.
Detection A	TBS3060C	220 µL of Biotin-mouse KC antibody (10).	May be stored for up to 3 months at 2-8 °C.*
Detection B	TBS3060D	300 µL of Streptavidin-HRP (50x).	
Assay Diluent	TBS3060E	15 mL of a buffered protein base with preservatives.	
Wash Buffer	TBS3000W	15 mL of concentrated solution (10x).	
TMB Substrate	TBS3000T	12 mL of ultra-sensitive TMB substrate.	
Stop Solution	TBS3000S	6 mL of 2 N sulfuric acid.	

Store the unopened kit at 2-8 °C. Do not use past kit expiration date. The kit contains sufficient materials to run an ELISA on one 96 well plate.

PRECAUTIONS

Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer: Add 12 mL of Wash Buffer Concentrate (10X) to 108 mL of deionized distilled water to prepare 120 mL of Wash Buffer (*If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved.*).

Detection A working solution preparation: Add 210 μL of **Detection A Stock** (Biotin-mouse KC antibody) to 1890 μL Assay Diluent to prepare Detection A working solution. Add 20 μL to each well.


Detection B working solution preparation: Add 240 μL of **Detection B** streptavidin-HRP to 12 mL Assay Diluent to prepare Detection B working solution.

Mouse KC Standard Preparation:

Label test tubes as #1 through #8. Pipet 490 μL of 1x Assay Diluent into tube #1, and 250 μL into tubes #2 to #8 as Fig.2 diagram below.

1. Add 10 μL of the Mouse KC Standard stock solution (50 ng/mL) by dilution of 50X to tube #1 and mix.
2. Make 2x serial dilutions using the of 1000 pg/mL (tube #1) standard solution from tube #2 through #7 with sequential transfer of 250 μL to the next concentration. Mix each tube thoroughly before the next transfer. The standard concentration in tube 1 through 7 will be 1000, 500, 250, 125, 62.5, 31.25, and 15.63 pg/mL. Tube# 8 is blank (0 pg/mL)

Fig.2 Diagram for Mouse KC standard preparation



ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended that all standards, controls, and samples be assayed in duplicate.

1. Add 80 μL of standard, sample, or control per well.
2. Add 20 μL of **Detection A** to the above standard and sample of each well, thoroughly mix. Cover with the adhesive sealer. Incubate at **RT for 2 hours with shaking**.
3. Aspirate each well (no wash). Invert the plate and blot it against clean paper towels.
4. Add 100 μL of **Detection B working solution** to each well. Incubate at **RT for 1 hour with shaking**.
5. Aspirate each well, and wash for 3 times by filling each well with 200 μL Wash Buffer (*Complete removal of liquid at each step is essential to good performance*). After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

6. Add 100 µL of **TMB Substrate** to each well. Incubate **at RT for 10-20 minutes with shaking** (*Protect from light*). The color becomes blue.
7. Add 50 µL of **Stop Solution** to each well. The color in the well should change from blue to yellow (gently tap the plate to ensure thorough mixing).
8. Determine the optical density of each well within 20 minutes, using a microplate reader at 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. If wavelength correction is not available, subtract readings at 540 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, control, and sample subtract the average zero standard optical density (O.D.).

Create a standard curve using computer software capable of generating a four-parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the Y-axis against the concentration on the X-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the mouse KC concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

TYPICAL DATA

This standard curve ($R^2=0.9998$) is provided for demonstration only. A standard curve should be generated for each set of samples assayed. Fig. 3 is an example of typical Data.

SENSITIVITY

The minimum detectable dose (MDD) of mouse KC is typically 10 pg/ml.

The Intra-assay CV is 3.79% the Inter-assay CV is <10%.


SPECIFICITY

This assay recognizes natural and recombinant mouse KC.

No cross-reactivity with any other cytokine is observed.

RELATIVE PRODUCTS

- TBS3030 Fast Mouse IL-1β ELISA
- TBS3031 Fast Mouse IL-2 ELISA
- TBS3032 Fast Mouse IL-4 ELISA
- TBS3040 Fast Mouse IL-6 ELISA
- TBS3044 Fast Mouse IL-10 ELISA
- TBS3047 Fast Mouse IL-12 p70 ELISA
- TBS3049 Fast Mouse IL-13 ELISA
- TBS3050 Fast Mouse TNF-α ELISA
- TBS3070 Fast Mouse NGF ELISA
- TBS3079 Fast Mouse GM-CSF ELISA
- TBS3080 Fast Mouse G-CSF ELISA
- TBS3084 Fast Mouse IFN-γ ELISA
- TBS3085 Fast Mouse TGF ELISA
- TBS3086 Fast Mouse MCPT-1 ELISA
- TBS3090 Fast Mouse IL-17AF ELISA
- TBS3091 Fast Mouse IL-19 ELISA
- TBS3092 Fast Mouse IL-21 ELISA
- TBS3093 Fast Mouse IL-22 ELISA
- TBS3094 Fast Mouse IL-23 ELISA
- TBS3095 Fast Mouse IL-27 ELISA
- TBS3096 Fast Mouse IL-28B ELISA
- TBS3097 Fast Mouse IL-33 ELISA



For research use only. Not for use in diagnostic procedures.