

Catalog Number	Kit Size
TBS42021-100	100 assays
TBS42021-200	200 assays

DESCRIPTION

The Aspergillus Flavus qPCR Kit has been designed to identify the aspergillus species of A Flavus in a single PCR reaction using real-time quantitative polymerase chain reaction(qPCR) and probe fluorescence labels. The detection of target DNA confirms ingredient authenticity and prevents food fraud, ethical issues, or health concerns.

PRINCIPLE

Authenticating ingredients utilizes real-time PCR which is based on the amplification of a specific region of the relevant target genome. The amplified product is detected using target-specific fluorescent probes that bind to the amplified product. As the PCR product accumulates, there is an increased fluorescent signal from the bound probes. Monitoring the fluorescence intensities during the PCR run allows the detection of the accumulating PCR product in real time.

The Aspergillus Flavus qPCR Kit includes aspergillus Flavus target positive and negative controls, PCR internal controls labeled with HEX, a qPCR super mix, and the primer-probe mix in which the probe is labeled with FAM. These aid in a straightforward interpretation of the results (see the table "Summary of possible PCR outcomes").

KEY FEATURES

- ❖ High sensitivity and specificity for Flavus detection.
- ❖ High efficiency: the optimal systemic conditions for PCR amplifications.
- ❖ Streamlined protocol: Just add DNA Template and water.
- ❖ No cross reactivity with other species.

APPLICATIONS

Detect aspergillus-derived DNA in plant, cannabis, cannabis ingredients, grain, food, herbals, and animal feed.

KIT CONTENTS

Name	100RXN	200RXN
qPCR Super Mix (FL1)	0.8mL	1.6mL
Primer-Probe Mix (FL2)	0.6mL	1.2mL
Positive Control DNA (FL ⁺)	60μL	100μL
Negative Control DNA (FL ⁻)	60μL	100μL

The Flavus probe is labeled with **FAM**, and PCR internal control is labeled with **Hex**.

STORAGE CONDITION

The kit is shipped on ice and stored at -20°C for long-term storage. Shelf life of 12 months after receipt.

PCR PROTOCOL

1. Set up PCR reaction for each sample in 20 μL

Reaction Component	Volume (μL)
qPCR Super Mix (FL1)	7.0
Primer-probe Mix (FL2)	5.0
Nuclease-free Water	3.0
DNA sample	5.0
Final Volume	20μL

Internal control should be included as below: Positive Control (5μL DNA /reaction) Negative Control (5μL DNA/reaction)

2. Suggested PCR conditions

Step	Amplification	PCR	
		CYCLE (40 cycles)	
		Denature	Anneal/ Extend
Temperature	95°C	95°C	60°C
Time	2 min	15 sec	60 sec

DATA ANALYSIS

Positive Reaction: Sample Ct ≤ 37, and Positive, Negative and Blank controls are normal.

Negative Reaction: Sample Ct ≥ 38, and Positive, Negative and Blank controls are normal.

PCR internal control is positive in all samples, positive and negative controls. The positive response indicates a normal PCR amplification. Otherwise, the PCR reaction may be inhibited.

Repeat Reaction: If one of the control reactions is not normal, PCR reaction is failed, and should be repeated.

RELATIVE PRODUCTS

TBS6025: Microbial DNA Magnetic Extraction
TBS42025:4-In-1 Aspergillus qPCR
TBS42026: O157H7 E. Coli qPCR
TBS42027: STEC qPCR
TBS42028: Salmonella qPCR
TBS42029: STEC and Salmonella Multiple qPCR
TBS42030: Mycoplasma Detection qPCR
TBS42031: Listeria Monocytogen qPCR
TBS42032: Listeria Genus qPCR
TBS42033: Bacillus Cereus qPCR
TBS 42020: Universal Aspergillus qPCR
TBS42021: Aspergillus Flavus qPCR
TBS42022: Aspergillus Fumigatus qPCR
TBS42023: Aspergillus Niger qPCR
TBS42024: Aspergillus Terreus qPCR

For research use only.