

## Fast Human AD7c NTP ELISA

For the quantitative determination of human AD7c NTP concentrations in cell culture supernates, serum, and plasma.

### INTRODUCTION


AD7c-NTP, also called Alzheimer-associated neuronal thread protein is a member of Neuronal Thread Proteins (NTPs); it can be detected in increased concentration in cortical neurons, brain-tissue extracts, cerebrospinal fluid, and urine in the early course of AD neurodegeneration, and its level is proportional to the degree of dementia, which makes it a promising biomarker for AD.

Tribioscience’s Fast Human AD7c NTP ELISA is designed to quantitatively detect human AD7c NTP levels in serum, plasma, and other biological samples. The main feature is that **the kit uses our novel proprietary approaches to combine samples and detections into a one-step instead of the complicated traditional methods. It makes the assay simple, easy, accurate and fast. The Hands-on time can be within 2 hours, no need for 4-5 hours (Fig. 1). The detection range is from 15 to 1000 ng/mL.** The levels of human AD7c NTP samples are parallel to the standard curves obtained using the kit standards linearly. Therefore, the kit can be used to determine relative mass values for natural human AD7c NTP protein.

### PRINCIPLE OF THE ASSAY

This assay employs our novel proprietary sandwich enzyme immunoassay techniques (see Fig. 1). A monoclonal antibody specific for human AD7c NTP is pre-coated onto a microplate. Standards or samples and a biotin conjugated detection antibody are pipetted into the wells and concurrently incubated to form a sandwich complex in one step. Simply aspirate each well without washing and directly add Streptavidin-HRP into the complex. Following a wash, an **ultra-sensitive TMB substrate solution** is added to the wells for color development. The color intensity is proportional to the amount of AD7c NTP bound in the initial step. The intensity of the color is measured by plate reading at 450 nm.

Fig. 1



### KIT CONTENT AND STORAGE CONDITIONS

PART	PART#	DESCRIPTION	STORAGE OF OPENED/ RECONSTITUTED
Human AD7c NTP Microplate	TBS3297A	96-well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody specific for human AD7c NTP.	Return unused wells to the foil pouch. Reseal along the entire edge of the zip-seal. May be stored for up to 1 month at 2-8 °C.
Human AD7c NTP Standard	TBS3297B	50 µl of Recombinant human AD7c NTP (10 µg/mL).	Aliquot and store at -20 °C for up to 1 month in a manual defrost freezer. Avoid repeated freeze-thaw cycles.
Detection A	TBS3297C	2.1 ml of biotin- human AD7c NTP antibody.	May be stored for up to 3 months at 2-8 °C.
Detection B	TBS3297D	12 ml of streptavidin HRP.	
Assay Diluent	TBS3297E	12 ml of a buffered protein base with preservatives.	
Wash Buffer	TBS3000W	12 ml of concentrated solution(10x).	
TMB Substrate	TBS3000T	12 ml of ultra-sensitive TMB substrate.	
Stop Solution	TBS3000S	6 ml of 2 N sulfuric acid.	

Store the unopened kit at 2-8 °C. Do not use past kit expiration date.

The kit contains sufficient materials to run an ELISA on one 96-well plate.

### PRECAUTIONS

Wear protective gloves, clothing, eye, and face protection. Wash hands thoroughly after handling.

**REAGENT PREPARATION**


Bring all reagents to room temperature before use.

**Wash Buffer:** Add 12 mL of Wash Buffer Concentrate (10x) to 108 mL of deionized distilled water to prepare 120 mL of Wash Buffer (If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved).

**Human AD7c NTP Standard Preparation:** Label test tubes as #1 through #8. Pipet 360 µL of 1x Assay Diluent into tube #1, and 200 µL into tubes #2 to #8 as diagram below.

1. Add 40 µL of the Human AD7c NTP Standard stock solution (10 µg/mL) to tube #1 and mix.
2. Make 2x serial dilutions of the standard using the Tube#1(1000 ng/mL standard solution) from Tube #2 through #7 with sequential transfer of 200 µL to the next concentration. Mix each tube thoroughly before the next transfer. The standard concentration in tube 1 through 7 will be 1000, 500, 250, 125, 62.5, 31.25 and 15.625 ng/mL. Tube# 8 is Standard 0.

**Fig.2 Diagram for Human AD7c NTP standard preparation**



**ASSAY PROCEDURE**

Bring all reagents and samples to room temperature before use. It is recommended that all standards, controls, and samples be assayed in duplicate.

1. Add 80 µL of standard, sample, or control per well.
2. Add 20 µL of **Detection A** to the above standard and sample of each well, thoroughly mix. Cover with the adhesive sealer. Incubate at **RT for 2 hours**.
3. Aspirate each well (no wash). Invert the plate and blot it against clean paper towels.
4. Add 100 µL of **Detection B** to each well. Incubate at **RT for 1 hour**.
5. Aspirate each well, and wash for 3 times by filling each well with 300 µL Wash Buffer (Complete removal of liquid at each step is essential to good performance). After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 µL of **TMB Substrate** to each well. Incubate at **RT for 10-20min** (Protect from light). The color becomes blue.
7. Add 50 µL of **Stop Solution** to each well. The color in the well should change from blue to yellow (gently tap the plate to ensure thorough mixing).
8. Determine the optical density of each well within 20 minutes, using a microplate reader at 450 nm. If wavelength correction is available, set to 542 nm or 570 nm. If wavelength correction is not available, subtract readings at 542 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.

**CALCULATION OF RESULTS**

Average the duplicate readings for each standard, control, and sample subtract the average zero standard optical density (O.D.).

Create a standard curve using computer software capable of generating a four-parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the Y-axis against the concentration on the X-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the human concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

**TYPICAL DATA**

This standard curve ( $R^2=1.000$ ) is provided for demonstration only. A standard curve should be generated for each set of samples assayed. Fig. 3 is an example of typical Data.

**SENSITIVITY**

The minimum detectable dose (MOD) of human is typically 10 ng/ml.

The Intra-assay CV and the Inter-assay CV are <10%.

**SPECIFICITY**


This assay recognizes natural and recombinant human AD7c NTP.

No cross-reactivity with others.

**RELATIVE PRODUCTS**

- Human p-Tau-217 ELISA (TBS3293)
- Human p-Tau-181 ELISA (TBS3294)
- Human Total Tau ELISA (TBS3295)
- Human p-Tau-231 ELISA (TBS3296)
- Human Amyloid  $\beta$ 42 ELISA (TBS3299)
- Human Amyloid  $\beta$ 40 ELISA (TBS3298)
- Human NF-L ELISA (TBS32101)
- Human Total Amyloid  $\beta$  ELISA (TBS32104)
- Human Gamma H2AX ELISA (TBS3265)
- Human H2AX ELISA (TBS3266)
- Human IL-4 ELISA (TBS3221)
- Human IL-4 ELISA (TBS3221)
- Human IL-6 ELISA (TBS3223)
- Human IL-7 ELISA (TBS3224)
- Human IL-8 ELISA (TBS3225)
- Human IL-10 ELISA (TBS3226)
- Human IL-13 ELISA (TBS3227)
- Human IL-17 ELISA (TBS3228)
- Human IL-22 ELISA (TBS3229)
- Human IL-33 ELISA (TBS4245)
- Human IFN-gamma ELISA (TBS3230)
- Human TGF-  $\beta$ 1 ELISA (TBS3232)
- Human GM-CSF ELISA (TBS3233)
- Human MIP-1 $\alpha$  ELISA (TBS3234)
- Protein Cell Lysis Buffer (catalog# TBS5001)
- Protein Assay Kit (Catalog# TBS2005)
- TMB Substrate System (Catalog#TBS5021)

**Fig.3 Human AD7c Standard Curve**



**For research use only. Not for use in diagnostic procedures.**